

No. D/21013/3282/14.12.2018/P-1  
GOVERNMENT OF INDIA  
MINISTRY OF HOME AFFAIRS  
P M DIVISION (PROCUREMENT WING)  
JAISALMER HOUSE, 26 MAN SINGH ROAD,  
NEW DELHI – 110 011  
TEL NO. 91-011-23387306 FAX NO. 91-011-23386764

Dated, the 8<sup>th</sup> May'2019

**OFFICE MEMORANDUM**

Global Tender Enquiry (GTE) No. D/21013/3282/14.12.2018/P-1 dated 08.01.2019 floated on CPP Portal and opened on 26.02.2019 (Tech bid) for procurement of Comparison Microscope – Qty. 05 Units for CFSLs (i) Bhopal (ii) Chandigarh (iii) Guwahati (iv) Kolkata and (v) Pune.

A GTE dated 08.01.2019 was floated on CPP portal and was opened on 26.02.2019 (Technical bid). The following four firms was participated in tender details as under :-

S.No.	Name of Firm
1	M/s Projectina AG Pvt. Ltd.
2	M/s KAV Engineering and Marketing Pvt. Ltd.
3	M/s Radical Scientific Equipment's Pvt. Ltd.
4	M/s DHR Holding India Pvt. Ltd

2. Tender Evaluation Committee (TEC) under the Chairmanship of Director, CFSL Hyderabad, carried out technical evaluation of technical bids of offers given by the above 04 firms against subject store mentioned in the said tender enquiry. The TEC thoroughly examined the qualification criteria, eligibility criteria and evaluated the techno-commercial and technical parameters of all the above bids. The TEC in its report mentioned few short comings/observations in the Technical Bid submitted by the following firms :-

(A) **M/s Projectina AG Pvt. Ltd.**

(i) Appendix-2 (Technical Specifications) Sl.No.2.2 at page 37 of GTE.

*"An optical comparison bridge with motorized Z-movement for height adjustment of suitable range"*

- The firm mentioned the static comparison bridge however, the user's requirement is for motorized comparison bridge with z- movement. The same was also telephonically clarified from OEM i.e. M/s Projectina AG Pvt. Ltd. and the firm confirmed that the Comparison Bridge is static.

Hence, Bidder is **ineligible**.

R

(B) **M/s Radical Scientific Equipment's Pvt. Ltd.**

(i) Appendix-2 (Technical Specifications) Sl.No.2.2 at page 37 of GTE.

*"An optical comparison bridge with motorized Z-movement for height adjustment of suitable range"*

- The firm does not mentioned regarding the movement of the comparison bridge and failed to clarify regarding the z movement of the comparison bridge.

(ii) Appendix-2 (Technical Specifications) Sl.No.2.3 at page 37 of GTE.

*"Suitable mechanical stages for left side and right side preferably with a digital scale or micrometer in X and Y directions. Coarse and fine focus adjustment for the left and right specimen stages with independent controls"*

- The firm has not supplied the details regarding left and right movement of the mechanical stages and digital / micrometer scales. Hence, the firm failed to clarify the same.

(iii) Appendix-2 (Technical Specifications) Sl.No.2.4 at page 37 of GTE.

*"Mechanical and/or motorized coaxial drives for X/Y with a range of movement of the stages of X ~ 50mm/ Y ~ 50 mm or more and synchronized movement of stages in X-direction"*

- The firm did not mentioned regarding the synchronized movement of both the stages. Hence, the firm failed to clarify the same.

(iv) Appendix-2 (Technical Specifications) Sl.No.2.5 at page 37 of GTE.

*"Suitable stage attachment with 360° rotating basic mount in horizontal plane and  $\pm 180^\circ$  angle function with scale divisions"*

- The firm failed to provide Sufficient clarification.

(v) Appendix-2 (Technical Specifications) Sl.No.2.9 at page 37 of GTE.

*"The comparison microscope bridge must offer the following observation modes: side-by-side view, full left or full right view, and superimposed view"*

- The Firm failed to clarify regarding the superimposition of two images.

(vi) Appendix-2 (Technical Specifications) Sl.No.2.10 at page 37 of GTE.

*"Provision for simultaneous binocular and video/photo observation/documentation or to allow full light to be viewed on binocular or full light to divert images to a camera."*

- The Firm failed to clarify whether full light to be viewed on binocular or full light to divert images to camera.

(vii) **Annexure II Clause 5 (i) at page 12 of GTE**

***"Tenderer should have supplied and installed successfully at least 1 equipment in the last 5 years before the date of opening of tender and functioning satisfactorily, without adverse report. Report from the institution to whom it has been supplied about satisfactory performance must be enclosed with technical bid."***

R

- The firm failed to provide the PO and Satisfactory Reports for the successful installation of Comparison Microscope. The PO provided by the firm with P.O No. FSL/Stores/256/2016 dated 28.11.2016 was confirmed from FSL Sagar and found that no such equipment i.e. Comparison Microscope of model no. RCM-12 was supplied by the firm to them

Hence, Bidder is **ineligible**.

**(C) M/s KAV Engineering and Marketing Pvt. Ltd**

- (i) Appendix-2 (Technical Specifications) Sl.No.3.3 at page 38 of GTE

*"The light sources must have all fixtures, holders, spare lamps, fuses, connecting cables, short and long articulated arms for the fiber-optic light guides, or any other item required for the illumination systems installation and working"*

- The firm denied to offer any spare lamp alongwith the equipment as per the required tender specifications.

- (ii) Appendix-2 (Technical Specifications) Sl.No.7.1 at page 38 of GTE

*"The firm must supply spares for a period of at least five (5) years and availability of spares must be ensured a period of at least for ten (10) years after the date of installation. List of spares and consumables with cost must be supplied."*

- The Firm denied to offer spares for 5 years as per the requirement of tender specifications.

Hence, Bidder is **in-eligible**.

3. Based on the observation/short coming noticed TEC recommended the following in its report :-

S.NO.	Name of Firm	Technically Qualified / Not Qualified
1	M/s Projectina AG Pvt. Ltd.	"NOT QUALIFIED"
2	M/s KAV Engineering and Marketing Pvt. Ltd.	"NOT QUALIFIED"
3	M/s Radical Scientific Equipment's Pvt. Ltd.	"NOT QUALIFIED"
4	M/s DHR Holding India Pvt. Ltd	"QUALIFIED"

4. This has the approval of AS(PM).

*R.D. Shivsharan*  
**(R.D SHIVSHARAN)**  
**DDG (Proc), MHA**

Copy to :-

1. The Director, DFSS, CFSL, Block No.9, CGO Complex, Lodhi Road, New Delhi.
2. The Section Officer (IT), IT Cell Website Room No.10 MHA, North Block New Delhi with request to get is incorporated in the official website of the MHA.
3. CPP Portal.